
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000194
8:48 18/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:52
STERILIZING 2:46
136.4 C 230kPa 2:46
136.3 C 232kPa 3:16
136.2 C 231kPa 3:46
136.4 C 233kPa 4:16
136.4 C 233kPa 4:46
136.3 C 230kPa 5:16
136.1 C 229kPa 5:46
136.4 C 230kPa 6:16
136.4 C 230kPa 6:16
VENTING 6:16
AIR DRYING 7:04
CYCLE COMPLETE 7:09

Digital Signature #
871B473CBE4E8846
