

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2471

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2023 11:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.8 F 0.0  
2:00 183.5 F 3.6  
4:00 198.0 F 9.0  
6:00 220.7 F 8.4  
8:00 230.6 F 10.5  
10:00 232.2 F 7.8  
12:00 241.1 F 11.1  
14:00 246.0 F 13.8  
16:00 258.5 F 20.6  
18:00 269.6 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.5 F 29.7  
2:00 271.2 F 29.0  
3:00 272.3 F 29.9  
4:00 272.3 F 29.5  
  
Min 271.0 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 220.9 F 0.0  
10:00 222.0 F 0.0  
15:00 216.5 F 0.0  
20:00 229.2 F 0.0  
25:00 226.5 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 18:59  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:14  
CYCLE COMPLETE

06/06/2023 12:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_