
Midmark M11 - v1.0.5
Total Cycles: 1924

Name: RITTER2
Sterilizer ID: V2313228
01/11/2023 05:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 192.9 F 4.0
4:00 207.7 F 10.0
6:00 224.7 F 6.4
8:00 239.0 F 12.6
10:00 240.2 F 11.2
12:00 239.9 F 10.6
14:00 254.7 F 18.0
16:00 267.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.2
2:00 271.4 F 28.6
3:00 272.9 F 29.4
4:00 272.0 F 28.8

Min 271.4 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 238.5 F 0.0
10:00 227.1 F 0.0
15:00 223.2 F 0.0
20:00 221.0 F 0.0
25:00 220.6 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 2:50
HEATING: 17:23
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:57:17
CYCLE COMPLETE

01/11/2023 06:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____