
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001750
11:27 19/04/2024
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 4:13
STERILIZING 5:09
136.4 C 223kPa 5:09
136.1 C 222kPa 5:39
136.4 C 225kPa 6:09
136.4 C 223kPa 6:39
136.4 C 224kPa 7:09
136.3 C 223kPa 7:39
136.4 C 222kPa 8:09
136.4 C 224kPa 8:39
136.4 C 225kPa 9:09
136.2 C 222kPa 9:39
136.1 C 221kPa 10:09
136.3 C 224kPa 10:39
136.3 C 224kPa 11:09
136.4 C 225kPa 11:39
136.4 C 223kPa 12:09
136.3 C 224kPa 12:39
136.0 C 221kPa 13:09
136.4 C 222kPa 13:39
136.4 C 225kPa 14:09
136.4 C 224kPa 14:39
136.4 C 223kPa 15:09
136.4 C 223kPa 15:09
VENTING 15:09
AIR DRYING 15:57
CYCLE COMPLETE 45:57

Digital Signature #
68656A7D85725270
