
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000731
7:13 30/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:59
STERILIZING 4:02
136.4 C 223kPa 4:02
136.3 C 222kPa 4:32
136.2 C 220kPa 5:02
136.4 C 222kPa 5:32
136.4 C 222kPa 6:02
136.1 C 222kPa 6:32
136.4 C 222kPa 7:02
136.4 C 224kPa 7:32
136.4 C 224kPa 7:32
VENTING 7:32
AIR DRYING 8:21
CYCLE COMPLETE 8:24

Digital Signature #
403122A3E688A092
