
Midmark M11 - v1.0.0
Total Cycles: 2886

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2023 06:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.2 F 0.0
2:00 168.3 F 2.9
4:00 175.0 F 6.8
6:00 196.3 F 11.6
8:00 211.0 F 5.0
10:00 225.0 F 9.9
12:00 235.6 F 8.3
14:00 239.8 F 11.9
16:00 239.0 F 10.1
18:00 251.9 F 16.8
20:00 263.7 F 23.7
22:00 271.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.6 F 29.6
2:00 271.8 F 28.9
3:00 272.5 F 29.8
4:00 272.3 F 29.3

Min 271.5 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 212.4 F 0.0
10:00 218.1 F 0.0
15:00 224.3 F 0.0
20:00 232.0 F 0.0
25:00 228.7 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:05
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

02/09/2023 07:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____