
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001404
10:47 09/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 5:16
STERILIZING 6:36
136.0 C 220kPa 6:36
136.0 C 220kPa 7:06
136.4 C 223kPa 7:36
136.5 C 224kPa 8:06
136.2 C 222kPa 8:36
136.4 C 221kPa 9:06
136.4 C 219kPa 9:36
136.4 C 224kPa 10:06
VENTING 10:06
AIR DRYING 11:05
CYCLE COMPLETE 11:11

Digital Signature #
05DA0CCFE425F838
