

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000453  
12:44 01/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:37  
PRESSURIZING 3:39  
STERILIZING 4:45  
136.5 C 233kPa 4:45  
136.3 C 232kPa 5:15  
135.9 C 230kPa 5:45  
136.4 C 233kPa 6:15  
136.4 C 231kPa 6:45  
136.3 C 232kPa 7:15  
136.2 C 232kPa 7:45  
136.3 C 232kPa 8:15  
136.3 C 232kPa 8:15  
VENTING 8:15  
AIR DRYING 9:10  
CYCLE COMPLETE 9:10  
-----

Digital Signature #  
3671285EDAB5637C  
-----