
Midmark M11 - v1.0.5
Total Cycles: 935

Name: RITTER2
Sterilizer ID: V2313228
03/31/2022 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 199.7 F 1.8
4:00 204.6 F 7.3
6:00 223.7 F 13.0
8:00 231.0 F 9.7
10:00 235.6 F 8.9
12:00 240.7 F 11.2
14:00 246.9 F 14.4
16:00 259.8 F 21.6
18:00 271.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.8 F 29.9
2:00 271.5 F 29.1
3:00 272.5 F 30.1
4:00 272.1 F 29.5

Min 271.3 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 213.8 F 0.0
10:00 234.8 F 0.0
15:00 236.7 F 0.0
20:00 227.4 F 0.0
25:00 226.6 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:39
CYCLE COMPLETE

03/31/2022 09:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____