
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002395
11:50 18/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:52
STERILIZING 4:03
136.4 C 224kPa 4:03
136.2 C 223kPa 4:33
135.9 C 222kPa 5:03
136.4 C 219kPa 5:33
136.4 C 223kPa 6:03
136.4 C 220kPa 6:33
135.9 C 221kPa 7:03
136.4 C 223kPa 7:33
136.4 C 223kPa 7:33
VENTING 7:33
AIR DRYING 8:23
CYCLE COMPLETE 12:13

Digital Signature #
7EEAFF34E77C1B09
