
Midmark M11 - v1.0.0
Total Cycles: 730

Name: RITTER 1
Sterilizer ID: V2178198
03/18/2021 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.5 F 0.0
2:00 141.5 F 1.5
4:00 162.2 F 4.5
6:00 189.0 F 9.1
8:00 221.7 F 7.0
10:00 220.8 F 8.3
12:00 237.0 F 11.9
14:00 236.9 F 11.1
16:00 237.5 F 9.1
18:00 251.8 F 16.0
20:00 264.0 F 23.0
22:00 270.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.7
2:00 272.5 F 29.5
3:00 272.1 F 29.0
4:00 271.3 F 28.7

Min 271.2 F 28.6
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 236.3 F 0.0
10:00 226.9 F 0.0
15:00 229.4 F 0.0
20:00 223.6 F 0.0
25:00 223.9 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:53
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

03/18/2021 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____