
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001816
10:22 18/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:41
STERILIZING 3:47
136.5 C 224kPa 3:47
136.3 C 224kPa 4:17
136.1 C 222kPa 4:47
136.4 C 225kPa 5:17
136.2 C 222kPa 5:47
136.2 C 222kPa 6:17
136.3 C 224kPa 6:47
136.4 C 224kPa 7:17
136.4 C 224kPa 7:17
VENTING 7:17
AIR DRYING 8:10
CYCLE COMPLETE 8:19

Digital Signature #
C52E1DF771D1AB89
