
Midmark M11 - v1.0.0
Total Cycles: 5052

Name: RITTER 1
Sterilizer ID: V2178198
11/04/2024 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.7 F 0.0
2:00 151.4 F 0.3
4:00 176.9 F 3.2
6:00 199.5 F 7.7
8:00 214.6 F 12.6
10:00 225.2 F 7.4
12:00 236.4 F 13.0
14:00 239.0 F 10.6
16:00 236.6 F 9.0
18:00 251.1 F 15.9
20:00 262.5 F 22.6
22:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.2
2:00 272.7 F 29.0
3:00 271.9 F 28.5
4:00 271.3 F 28.4

Min 271.3 F 28.2
Max 273.2 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.3 F 0.0
5:00 198.2 F 0.0
10:00 198.0 F 0.0
15:00 223.2 F 0.0
20:00 237.7 F 0.0
25:00 232.0 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

11/04/2024 03:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____