
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007400
8:25 21/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:23
STERILIZING 3:28
136.3 C 220kPa 3:28
136.4 C 223kPa 3:58
136.4 C 224kPa 4:28
136.4 C 223kPa 4:58
136.3 C 224kPa 5:28
136.3 C 222kPa 5:58
136.4 C 222kPa 6:28
136.4 C 224kPa 6:58
136.4 C 224kPa 6:58
VENTING 6:58
AIR DRYING 7:49
CYCLE COMPLETE 8:03

Digital Signature #
BE951AC23F533B4C
