

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3659

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2023 04:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 216.6 F 0.0  
2:00 212.4 F 4.6  
4:00 210.6 F 10.1  
6:00 222.2 F 6.1  
8:00 235.3 F 11.0  
10:00 234.4 F 8.5  
12:00 240.1 F 10.2  
14:00 247.6 F 14.0  
16:00 260.4 F 21.1  
18:00 271.8 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 271.6 F 28.2  
2:00 271.7 F 28.2  
3:00 271.9 F 28.4  
4:00 272.2 F 28.7  
  
Min 271.5 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 231.8 F 0.0  
10:00 230.0 F 0.0  
15:00 228.7 F 0.0  
20:00 226.3 F 0.0  
25:00 225.1 F 0.0  
30:00 221.7 F 0.0  
Drying Complete

FILLING: 2:33  
HEATING: 18:27  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

08/15/2023 05:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_