

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3235

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2023 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.9 F 0.0  
2:00 77.4 F 0.0  
4:00 88.4 F 0.2  
6:00 120.1 F 0.9  
8:00 153.7 F 3.2  
10:00 174.9 F 6.5  
12:00 189.8 F 9.8  
14:00 216.1 F 13.3  
16:00 220.6 F 6.1  
18:00 232.2 F 9.9  
20:00 236.0 F 10.4  
22:00 238.7 F 10.4  
24:00 236.7 F 8.7  
26:00 247.6 F 14.1  
28:00 258.4 F 20.1  
30:00 268.4 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 272.7 F 29.5  
2:00 272.5 F 29.6  
3:00 271.8 F 29.4  
4:00 271.6 F 29.0

Min 270.9 F 28.5  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 235.9 F 0.0  
10:00 230.3 F 0.0  
15:00 217.8 F 0.0  
20:00 209.6 F 0.0  
25:00 207.2 F 0.0  
30:00 206.0 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:08  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:09:23  
CYCLE COMPLETE

12/20/2023 11:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_