
Midmark M11 - v1.0.0
Total Cycles: 2073

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.7 F 0.0
2:00 166.7 F 2.7
4:00 174.6 F 6.3
6:00 196.5 F 11.0
8:00 215.4 F 4.2
10:00 225.9 F 9.5
12:00 235.0 F 9.6
14:00 239.3 F 11.2
16:00 236.8 F 9.1
18:00 251.0 F 15.9
20:00 262.8 F 22.7
22:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.2
2:00 272.0 F 28.5
3:00 271.7 F 28.6
4:00 272.7 F 29.1

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 201.3 F 0.0
10:00 219.5 F 0.0
15:00 226.6 F 0.0
20:00 225.6 F 0.0
25:00 227.6 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

06/29/2022 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____