

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3307

Name: RITTER 1  
Sterilizer ID: V2178198  
05/30/2023 02:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.3 F 0.0  
2:00 192.6 F 2.4  
4:00 215.5 F 7.9  
6:00 225.7 F 12.3  
8:00 232.9 F 9.5  
10:00 235.9 F 8.9  
12:00 245.9 F 13.1  
14:00 245.1 F 12.8  
16:00 258.4 F 19.7  
18:00 269.8 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.6 F 28.9  
2:00 271.7 F 28.1  
3:00 272.3 F 28.8  
4:00 272.4 F 28.7  
  
Min 271.3 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 224.5 F 0.0  
10:00 227.8 F 0.0  
15:00 226.8 F 0.0  
20:00 223.3 F 0.0  
25:00 221.7 F 0.0  
30:00 218.1 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 18:57  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:16  
CYCLE COMPLETE

05/30/2023 03:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_