

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1567

Name: RITTER 1  
Sterilizer ID: V2178198  
01/26/2022 01:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.2 F 0.0  
2:00 190.9 F 4.2  
4:00 195.3 F 9.0  
6:00 216.4 F 7.5  
8:00 226.2 F 8.5  
10:00 238.9 F 12.3  
12:00 239.7 F 10.9  
14:00 237.8 F 8.9  
16:00 250.8 F 15.5  
18:00 262.3 F 22.1  
20:00 272.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.6 F 28.8  
2:00 271.8 F 28.1  
3:00 272.4 F 28.9  
4:00 272.3 F 28.6  
  
Min 271.5 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 227.8 F 0.0  
10:00 227.1 F 0.0  
15:00 230.0 F 0.0  
20:00 229.6 F 0.0  
25:00 228.2 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:22  
CYCLE COMPLETE

01/26/2022 02:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_