
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000404
9:04 26/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:31
STERILIZING 4:38
136.4 C 233kPa 4:38
136.4 C 233kPa 5:08
136.1 C 230kPa 5:38
136.2 C 232kPa 6:08
136.2 C 231kPa 6:38
136.4 C 231kPa 7:08
136.2 C 231kPa 7:38
136.5 C 233kPa 8:08
136.5 C 233kPa 8:08
VENTING 8:08
AIR DRYING 9:03
CYCLE COMPLETE 9:04

Digital Signature #
3E5927AF7BBBF974
