

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007996  
7:34 24/10/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:54  
PRESSURIZING 2:59  
STERILIZING 4:00  
136.0 C 220kPa 4:00  
136.4 C 223kPa 4:30  
136.4 C 221kPa 5:00  
136.4 C 219kPa 5:30  
136.2 C 220kPa 6:00  
136.0 C 219kPa 6:30  
136.1 C 219kPa 7:00  
136.3 C 220kPa 7:30  
136.3 C 220kPa 7:30  
VENTING 7:31  
AIR DRYING 8:19  
CYCLE COMPLETE 38:19

-----

Digital Signature #  
C684055C8F375420  
-----