
Midmark M11 - v1.0.5
Total Cycles: 1090

Name: RITTER2
Sterilizer ID: V2313228
05/10/2022 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.8 F 0.0
2:00 129.0 F 1.4
4:00 149.2 F 4.3
6:00 173.6 F 8.2
8:00 195.6 F 12.8
10:00 212.7 F 6.2
12:00 226.6 F 10.9
14:00 228.5 F 6.9
16:00 240.6 F 12.3
18:00 239.2 F 10.4
20:00 251.7 F 16.6
22:00 262.5 F 22.9
24:00 272.6 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.5 F 29.7
2:00 271.9 F 29.0
3:00 271.5 F 29.3
4:00 272.5 F 29.6

Min 271.1 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 209.0 F 0.0
10:00 225.5 F 0.0
15:00 224.8 F 0.0
20:00 232.4 F 0.0
25:00 226.8 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:39
CYCLE COMPLETE

05/10/2022 02:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____