
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001217
10:24 19/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:40
STERILIZING 4:34
136.4 C 222kPa 4:34
136.2 C 223kPa 5:04
136.3 C 224kPa 5:34
136.2 C 220kPa 6:04
136.2 C 220kPa 6:34
136.4 C 223kPa 7:04
136.4 C 225kPa 7:34
136.4 C 225kPa 8:04
136.4 C 225kPa 8:04
VENTING 8:04
AIR DRYING 9:00
CYCLE COMPLETE 9:07

Digital Signature #
8D5F666D0E404583
