
Midmark M11 - v1.0.0
Total Cycles: 3428

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2023 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 195.4 F 1.3
4:00 197.9 F 4.9
6:00 208.5 F 9.7
8:00 222.1 F 7.3
10:00 233.2 F 10.2
12:00 233.8 F 7.9
14:00 244.9 F 12.6
16:00 244.9 F 12.5
18:00 257.5 F 19.2
20:00 268.6 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.3 F 28.9
2:00 272.0 F 28.4
3:00 272.0 F 28.7
4:00 272.4 F 28.7

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 221.1 F 0.0
10:00 233.8 F 0.0
15:00 224.3 F 0.0
20:00 221.9 F 0.0
25:00 219.5 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

06/22/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____