
Midmark M11 - v1.0.5
Total Cycles: 347

Name: RITTER2
Sterilizer ID: V2313228
09/09/2021 12:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.8 F 0.0
2:00 146.0 F 2.0
4:00 167.4 F 5.5
6:00 190.5 F 9.7
8:00 211.2 F 9.2
10:00 227.0 F 8.5
12:00 239.1 F 13.5
14:00 238.1 F 11.0
16:00 237.0 F 9.6
18:00 250.3 F 16.0
20:00 261.3 F 22.4
22:00 272.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.6 F 29.7
2:00 271.0 F 29.1
3:00 272.5 F 29.5
4:00 272.4 F 29.5

Min 271.0 F 29.0
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 217.6 F 0.0
10:00 228.3 F 0.0
15:00 214.3 F 0.0
20:00 210.5 F 0.0
25:00 224.0 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:02:00
CYCLE COMPLETE

09/09/2021 01:26 PM
ACCEPTED: X
REJECTED:
Operator: U02