

-----  
Midmark M11 - v1.0.0  
Total Cycles: 734

Name: RITTER 1  
Sterilizer ID: V2178198  
03/22/2021 06:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.7 F 0.0  
2:00 72.8 F 0.0  
4:00 83.0 F 0.0  
6:00 121.8 F 0.1  
8:00 165.1 F 0.2  
10:00 194.5 F 1.2  
12:00 224.5 F 4.8  
14:00 232.1 F 7.8  
16:00 240.3 F 11.4  
18:00 228.0 F 6.3  
20:00 244.0 F 12.1  
22:00 238.9 F 9.5  
24:00 236.8 F 8.6  
26:00 249.7 F 14.6  
28:00 261.7 F 21.3  
30:00 272.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.7  
1:00 272.1 F 27.9  
2:00 272.6 F 28.4  
3:00 271.5 F 28.0  
4:00 272.1 F 28.0

Min 271.3 F 27.6  
Max 272.6 F 28.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
5:00 230.0 F 0.0  
10:00 245.0 F 0.0  
15:00 229.4 F 0.0  
20:00 224.5 F 0.0  
Dry Cycle Aborted

FILLING: 0:49  
HEATING: 30:19  
STERILIZING: 4:00  
VENTING: 2:50  
DRYING: 23:40  
TOTAL CYCLE: 01:01:39

03/22/2021 07:46 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01