
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001284
7:28 26/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:31
STERILIZING 4:44
136.4 C 223kPa 4:44
136.4 C 222kPa 5:14
136.1 C 221kPa 5:44
136.3 C 223kPa 6:14
136.4 C 223kPa 6:44
136.1 C 221kPa 7:14
136.2 C 222kPa 7:44
136.4 C 223kPa 8:14
136.4 C 223kPa 8:14
VENTING 8:14
AIR DRYING 9:06
CYCLE COMPLETE 9:16

Digital Signature #
C8865AEB00D5D1BE
