
Midmark M11 - v1.0.5
Total Cycles: 4320

Name: RITTER2
Sterilizer ID: V2313228
01/30/2025 04:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 153.5 F 2.1
4:00 163.6 F 5.7
6:00 186.4 F 9.9
8:00 203.2 F 7.5
10:00 218.0 F 7.4
12:00 229.7 F 11.6
14:00 232.0 F 8.2
16:00 244.8 F 13.1
18:00 242.7 F 12.3
20:00 254.4 F 18.3
22:00 265.2 F 24.6
24:00 271.2 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.5
1:00 271.8 F 29.4
2:00 272.5 F 29.7
3:00 271.9 F 30.0
4:00 271.1 F 29.7

Min 270.9 F 29.4
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 232.8 F 0.0
10:00 229.7 F 0.0
15:00 205.7 F 0.0
20:00 178.9 F 0.0
25:00 191.9 F 0.0
30:00 180.9 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:03:39
CYCLE COMPLETE

01/30/2025 05:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____