
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000164
14:22 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:56
PRESSURIZING 3:57
STERILIZING 5:25
136.5 C 225kPa 5:25
136.2 C 224kPa 5:55
136.2 C 224kPa 6:25
136.4 C 225kPa 6:55
136.4 C 226kPa 7:25
136.2 C 224kPa 7:55
136.4 C 226kPa 8:25
136.3 C 224kPa 8:55
136.3 C 224kPa 8:55
VENTING 8:55
AIR DRYING 9:58
CYCLE COMPLETE 13:54

Digital Signature #
6DEA74A4030AABE8
