

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5272

Name: RITTER 1  
Sterilizer ID: V2178198  
01/23/2025 07:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.3 F 0.0  
2:00 169.9 F 4.0  
4:00 185.3 F 8.5  
6:00 203.3 F 9.5  
8:00 219.7 F 7.1  
10:00 233.8 F 12.2  
12:00 234.2 F 8.8  
14:00 239.8 F 10.0  
16:00 246.3 F 13.4  
18:00 258.9 F 20.2  
20:00 270.2 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.3  
1:00 272.7 F 29.1  
2:00 271.8 F 28.3  
3:00 272.6 F 28.8  
4:00 271.9 F 28.5  
  
Min 271.6 F 28.2  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 203.5 F 0.0  
10:00 207.2 F 0.0  
15:00 216.6 F 0.0  
20:00 222.1 F 0.0  
25:00 221.7 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:42  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 00:59:27  
CYCLE COMPLETE

01/23/2025 08:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_