
Midmark M11 - v1.0.0
Total Cycles: 2519

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2022 12:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 164.5 F 3.3
4:00 181.7 F 7.2
6:00 208.8 F 12.4
8:00 220.9 F 6.4
10:00 233.2 F 11.6
12:00 233.7 F 8.4
14:00 240.6 F 10.6
16:00 245.8 F 13.2
18:00 258.3 F 19.9
20:00 269.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 29.1
2:00 271.7 F 28.7
3:00 272.7 F 29.1
4:00 271.6 F 28.5

Min 271.5 F 28.4
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 226.9 F 0.0
10:00 239.2 F 0.0
15:00 241.7 F 0.0
20:00 223.7 F 0.0
25:00 217.8 F 0.0
30:00 214.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

11/07/2022 01:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____