
Midmark M11 - v1.0.0
Total Cycles: 2652

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2022 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.2 F 0.0
2:00 140.9 F 0.3
4:00 160.3 F 1.8
6:00 186.4 F 5.4
8:00 209.6 F 9.9
10:00 218.7 F 6.8
12:00 229.4 F 9.3
14:00 236.3 F 10.2
16:00 240.4 F 11.4
18:00 237.9 F 9.5
20:00 251.7 F 16.1
22:00 263.2 F 22.8
24:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.0
2:00 271.5 F 28.2
3:00 272.5 F 29.1
4:00 271.8 F 28.4

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 209.7 F 0.0
10:00 220.0 F 0.0
15:00 226.2 F 0.0
20:00 225.9 F 0.0
25:00 226.9 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:02:09
CYCLE COMPLETE

12/07/2022 01:25 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____