
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001015
15:50 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:07
STERILIZING 3:00
136.0 C 230kPa 3:00
136.3 C 231kPa 3:30
136.3 C 234kPa 4:00
136.1 C 230kPa 4:30
136.4 C 233kPa 5:00
136.3 C 233kPa 5:30
136.3 C 230kPa 6:00
136.4 C 233kPa 6:30
136.4 C 233kPa 6:30
VENTING 6:30
AIR DRYING 7:20
CYCLE COMPLETE 7:54

Digital Signature #
61C5E3156F833CD7
