
Midmark M11 - v1.0.0
Total Cycles: 4469

Name: RITTER 1
Sterilizer ID: V2178198
04/17/2024 04:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.0 F 0.0
2:00 122.3 F 1.6
4:00 140.4 F 4.1
6:00 165.6 F 7.3
8:00 185.8 F 10.9
10:00 208.5 F 4.6
12:00 214.6 F 7.1
14:00 224.5 F 11.1
16:00 227.8 F 6.6
18:00 239.0 F 11.6
20:00 236.6 F 9.4
22:00 249.2 F 15.2
24:00 259.9 F 21.1
26:00 269.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.4
2:00 272.0 F 28.9
3:00 271.3 F 28.6
4:00 272.4 F 29.4

Min 271.3 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 207.5 F 0.0
10:00 226.2 F 0.0
15:00 228.9 F 0.0
20:00 222.3 F 0.0
25:00 222.4 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 27:02
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:05:56
CYCLE COMPLETE

04/17/2024 05:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____