
Midmark M11 - v1.0.0
Total Cycles: 3560

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2023 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.3 F 0.0
2:00 131.9 F 0.0
4:00 134.9 F 0.2
6:00 165.5 F 2.1
8:00 194.8 F 4.7
10:00 202.2 F 8.5
12:00 216.1 F 12.4
14:00 224.4 F 6.7
16:00 235.9 F 11.2
18:00 232.5 F 8.0
20:00 245.0 F 12.8
22:00 243.6 F 12.3
24:00 255.3 F 18.3
26:00 265.3 F 24.2
28:00 271.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.2
2:00 272.2 F 28.8
3:00 271.2 F 28.4
4:00 272.3 F 29.2

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 214.6 F 0.0
10:00 230.1 F 0.0
15:00 229.3 F 0.0
20:00 228.1 F 0.0
25:00 228.9 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:04
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:06:25
CYCLE COMPLETE

07/21/2023 10:39 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____