
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 006686
15:27 04/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:40
STERILIZING 3:47
136.4 C 224kPa 3:47
136.4 C 225kPa 4:17
136.5 C 219kPa 4:47
136.0 C 224kPa 5:17
136.3 C 219kPa 5:47
136.4 C 225kPa 6:17
136.4 C 225kPa 6:47
136.4 C 224kPa 7:17
136.4 C 224kPa 7:17
VENTING 7:18
AIR DRYING 8:10
CYCLE COMPLETE 8:18

Digital Signature #
92C6E2545C276340
