
Midmark M11 - v1.0.5
Total Cycles: 2386

Name: RITTER2
Sterilizer ID: V2313228
05/18/2023 07:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.1 F 0.0
2:00 78.6 F 0.0
4:00 89.9 F 0.0
6:00 106.9 F 0.2
8:00 147.8 F 0.3
10:00 187.0 F 0.7
12:00 207.5 F 4.4
14:00 217.3 F 7.9
16:00 227.6 F 11.7
18:00 228.4 F 7.3
20:00 239.9 F 12.4
22:00 238.6 F 10.1
24:00 236.8 F 8.8
26:00 249.4 F 15.0
28:00 260.7 F 21.4
30:00 271.1 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.3
2:00 271.6 F 28.5
3:00 271.9 F 29.2
4:00 272.4 F 29.1

Min 271.2 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 199.5 F 0.0
10:00 209.7 F 0.0
15:00 219.5 F 0.0
20:00 224.3 F 0.0
25:00 222.6 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:45
CYCLE COMPLETE

05/18/2023 08:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____