
Midmark M11 - v1.0.0
Total Cycles: 3804

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2023 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.2 F 0.0
2:00 124.5 F 0.0
4:00 139.2 F 0.6
6:00 185.2 F 2.9
8:00 188.1 F 6.4
10:00 207.5 F 10.7
12:00 224.9 F 5.6
14:00 232.4 F 8.9
16:00 241.7 F 13.6
18:00 239.7 F 10.6
20:00 235.7 F 8.2
22:00 250.0 F 14.9
24:00 261.5 F 21.4
26:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.3
1:00 272.6 F 28.7
2:00 272.6 F 28.9
3:00 271.7 F 28.2
4:00 272.3 F 28.6

Min 271.5 F 28.1
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 233.6 F 0.0
10:00 213.7 F 0.0
15:00 217.1 F 0.0
20:00 196.7 F 0.0
25:00 209.2 F 0.0
30:00 198.2 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 27:04
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:05:11
CYCLE COMPLETE

09/21/2023 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____