
Midmark M11 - v1.0.0
Total Cycles: 1053

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2021 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.2 F 0.0
2:00 198.7 F 4.1
4:00 211.4 F 9.9
6:00 224.5 F 7.2
8:00 235.5 F 12.0
10:00 236.2 F 9.8
12:00 237.7 F 8.6
14:00 249.6 F 15.2
16:00 261.5 F 22.0
18:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.2 F 28.9
2:00 271.7 F 28.8
3:00 272.2 F 28.9
4:00 271.4 F 28.6

Min 271.4 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.7 F 0.0
5:00 237.0 F 0.0
10:00 228.8 F 0.0
15:00 229.5 F 0.0
20:00 197.1 F 0.0
25:00 188.0 F 0.0
Dry Cycle Aborted

FILLING: 2:19
HEATING: 18:27
STERILIZING: 4:00
VENTING: 3:10
DRYING: 26:28
TOTAL CYCLE: 00:54:23

07/26/2021 12:20 PM
ACCEPTED: X
REJECTED:
Operator: U01