
Midmark M11 - v1.0.5
Total Cycles: 2770

Name: RITTER2
Sterilizer ID: V2313228
08/15/2023 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.6 F 0.0
2:00 190.8 F 0.0
4:00 207.3 F 1.6
6:00 223.0 F 6.6
8:00 235.7 F 11.7
10:00 235.8 F 9.4
12:00 238.1 F 9.5
14:00 241.3 F 10.8
16:00 248.8 F 14.8
18:00 261.5 F 22.0
20:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.2 F 28.8
2:00 272.8 F 29.3
3:00 272.1 F 29.1
4:00 272.1 F 28.7

Min 271.6 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 215.5 F 0.0
10:00 226.9 F 0.0
15:00 231.4 F 0.0
20:00 226.8 F 0.0
25:00 223.2 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

08/15/2023 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____