
Midmark M11 - v1.0.5
Total Cycles: 583

Name: RITTER2
Sterilizer ID: V2313228
11/29/2021 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.0 F 0.0
2:00 164.9 F 3.4
4:00 186.6 F 8.1
6:00 208.2 F 8.9
8:00 230.7 F 9.7
10:00 232.4 F 8.5
12:00 244.1 F 13.2
14:00 244.4 F 12.8
16:00 256.8 F 19.7
18:00 268.4 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 271.5 F 29.2
2:00 272.7 F 30.0
3:00 272.7 F 29.5
4:00 272.1 F 29.1

Min 271.1 F 28.8
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 221.0 F 0.0
10:00 218.1 F 0.0
15:00 224.2 F 0.0
20:00 221.8 F 0.0
25:00 225.8 F 0.0
30:00 228.3 F 0.0

Drying Complete

FILLING: 2:20
HEATING: 19:07
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

11/29/2021 03:18 PM
ACCEPTED: X
REJECTED:
Operator: U02