
Midmark M11 - v1.0.0
Total Cycles: 1288

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2021 08:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.4 F 0.0
2:00 149.2 F 1.1
4:00 161.6 F 3.8
6:00 191.5 F 8.4
8:00 212.9 F 8.2
10:00 223.6 F 7.8
12:00 236.5 F 12.7
14:00 239.0 F 9.9
16:00 236.2 F 8.4
18:00 249.7 F 14.8
20:00 261.6 F 21.7
22:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.9 F 29.3
2:00 272.4 F 28.9
3:00 271.9 F 28.4
4:00 271.8 F 28.7

Min 271.4 F 28.2
Max 273.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.5 F 0.0
5:00 217.1 F 0.0
10:00 189.7 F 0.0
15:00 186.0 F 0.0
20:00 182.9 F 0.0
25:00 180.5 F 0.0
30:00 178.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 23:18
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

10/22/2021 09:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____