
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005529
10:59 15/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:33
STERILIZING 3:26
136.5 C 223kPa 3:26
136.1 C 222kPa 3:56
136.4 C 225kPa 4:26
136.2 C 222kPa 4:56
136.1 C 221kPa 5:26
136.3 C 225kPa 5:56
136.4 C 224kPa 6:26
136.4 C 218kPa 6:56
136.4 C 218kPa 6:56
VENTING 6:56
AIR DRYING 7:49
CYCLE COMPLETE 10:00

Digital Signature #
E1FE79C31D41662B
