
Midmark M11 - v1.0.5
Total Cycles: 2931

Name: RITTER2
Sterilizer ID: V2313228
09/26/2023 01:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.9 F 0.0
2:00 179.1 F 3.3
4:00 191.9 F 7.7
6:00 231.8 F 12.8
8:00 229.2 F 7.7
10:00 240.6 F 12.6
12:00 236.4 F 10.1
14:00 236.9 F 8.9
16:00 249.7 F 15.3
18:00 261.0 F 21.8
20:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.3
1:00 273.2 F 29.6
2:00 272.7 F 29.9
3:00 271.5 F 29.3
4:00 272.4 F 29.5

Min 271.4 F 29.2
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 226.7 F 0.0
10:00 206.0 F 0.0
15:00 201.9 F 0.0
20:00 197.5 F 0.0
25:00 193.3 F 0.0
30:00 189.7 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

09/26/2023 02:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____