
Midmark M11 - v1.0.5
Total Cycles: 1257

Name: RITTER2
Sterilizer ID: V2313228
07/05/2022 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.2 F 0.0
2:00 132.1 F 1.6
4:00 147.2 F 4.9
6:00 172.0 F 9.2
8:00 193.1 F 9.0
10:00 215.2 F 7.7
12:00 229.7 F 12.7
14:00 232.1 F 9.4
16:00 238.8 F 10.0
18:00 244.6 F 13.6
20:00 256.6 F 19.9
22:00 267.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 271.8 F 29.7
2:00 272.5 F 30.1
3:00 271.7 F 29.9
4:00 272.0 F 29.7

Min 271.5 F 29.5
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 193.0 F 0.0
10:00 191.3 F 0.0
15:00 195.0 F 0.0
20:00 199.3 F 0.0
25:00 210.8 F 0.0
30:00 207.6 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:11
CYCLE COMPLETE

07/05/2022 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____