

-----  
Midmark M11 - v1.0.0  
Total Cycles: 935

Name: RITTER 1  
Sterilizer ID: V2178198  
06/24/2021 09:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.3 F 0.0  
2:00 160.7 F 1.5  
4:00 176.7 F 4.1  
6:00 195.8 F 7.4  
8:00 211.1 F 11.6  
10:00 223.5 F 5.2  
12:00 231.2 F 9.7  
14:00 235.0 F 9.8  
16:00 241.4 F 11.4  
18:00 238.4 F 9.4  
20:00 251.0 F 15.6  
22:00 261.8 F 21.7  
24:00 271.7 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 28.8  
2:00 271.3 F 28.1  
3:00 271.9 F 28.8  
4:00 272.5 F 28.7

Min 271.0 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
5:00 201.0 F 0.0  
10:00 198.4 F 0.0  
15:00 194.6 F 0.0  
20:00 193.1 F 0.0  
25:00 203.0 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 24:36  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:03:02  
CYCLE COMPLETE

06/24/2021 10:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_