
Midmark M11 - v1.0.0
Total Cycles: 1433

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2021 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 182.6 F 2.9
4:00 196.4 F 7.5
6:00 216.2 F 13.2
8:00 230.2 F 7.8
10:00 241.7 F 13.2
12:00 240.1 F 10.7
14:00 237.0 F 8.9
16:00 251.6 F 15.9
18:00 263.5 F 22.7
20:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.8 F 28.2
2:00 272.4 F 28.7
3:00 271.8 F 28.4
4:00 271.8 F 28.2

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 241.6 F 0.0
10:00 250.8 F 0.0
15:00 229.3 F 0.0
20:00 225.1 F 0.0
25:00 227.2 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

12/08/2021 01:42 PM
ACCEPTED: X
REJECTED:
Operator: U02