
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001835
10:52 12/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:00
STERILIZING 3:10
136.4 C 225kPa 3:10
136.4 C 226kPa 3:40
136.4 C 227kPa 4:10
136.2 C 224kPa 4:40
136.4 C 226kPa 5:10
136.4 C 225kPa 5:40
136.1 C 222kPa 6:10
136.3 C 225kPa 6:40
136.3 C 225kPa 6:40
VENTING 6:40
AIR DRYING 7:29
CYCLE COMPLETE 7:57

Digital Signature #
E20BAFFD6A81B195
