
Midmark M11 - v1.0.5
Total Cycles: 2545

Name: RITTER2
Sterilizer ID: V2313228
06/21/2023 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.4 F 0.0
2:00 175.7 F 3.6
4:00 181.1 F 7.8
6:00 205.2 F 13.1
8:00 221.8 F 7.3
10:00 234.5 F 12.5
12:00 236.2 F 9.9
14:00 236.7 F 8.8
16:00 249.1 F 15.2
18:00 261.1 F 21.9
20:00 271.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.7 F 29.7
2:00 272.0 F 29.0
3:00 271.8 F 29.3
4:00 272.7 F 29.7

Min 271.5 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 223.3 F 0.0
10:00 232.3 F 0.0
15:00 223.0 F 0.0
20:00 221.4 F 0.0
25:00 222.4 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

06/21/2023 11:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____