
Midmark M11 - v1.0.0
Total Cycles: 4396

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2024 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.9 F 0.0
2:00 171.5 F 2.2
4:00 173.2 F 5.6
6:00 193.1 F 9.3
8:00 209.3 F 13.2
10:00 219.5 F 6.5
12:00 228.2 F 10.7
14:00 230.5 F 7.2
16:00 241.2 F 12.0
18:00 239.4 F 10.1
20:00 251.5 F 16.0
22:00 261.9 F 21.9
24:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.0
2:00 272.2 F 28.6
3:00 271.2 F 28.1
4:00 272.0 F 28.8

Min 271.1 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 225.5 F 0.0
10:00 226.1 F 0.0
15:00 230.1 F 0.0
20:00 238.4 F 0.0
25:00 229.3 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 24:42
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:39
CYCLE COMPLETE

03/22/2024 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____